Hersey Measurement Company
Calibration Data

RF3
R&D

Date: 3-2-95
Order #: 545820
Serial #: 9304862
Calibrated by: 1222

Model 1050
(3 Wire)

Full Scale Flow Rate _____

Bridge Resistance _____

Full Scale mV/V _____

Calibration Resistor _____

Calibration Values:

Signal _____ mV/V

TP_5 (0-10 V) _____ V DC

Current Out _____ mA

Frequency Out _____ hZ

Model TW

Full Scale mV/V 2.0115

Bridge Resistance 350.6 ohms

Calibration Resistor (R7) 47.286 K ohms

Calibration Signal 1.8468 mV/V

Calibration Reading: 18.69 mA

•☒•☐•☐•☐• LO HI

•☐•☐•☐•☒•☐• - ZERO +

Additional Information _____

Model 1100
(Squirt)

Model TW

Full Scale mV/V _____

Bridge Resistance _____ ohms

Calibration Resistor (R7) _____ K ohms

Calibration Signal _____ mV/V

Calibration Reading: _____ mA

•☐•☐•☐•☐• LO HI

•☐•☐•☐•☐•☐• - ZERO +

Indicator/Totalizer

Digits: 3.5 4.5

Rate Decimal: 1 2 3 4

Total Decimal: 1 2 3 4 5 6 7 8

Scale: SEC MIN HRS

Function: LINEAR SQ. ROOT

Rate Low: _____

Rate Hi: _____

Cal. Low: _____ 4 mA= _____

Cal. Hi: _____ mA= _____

Additional Information: _____